Feasibility, Safety and Efficacy of Transcatheter Aortic Valve Replacement Without Balloon Pre-Dilation: A Systematic Review and Meta-Analysis

Vincent Auffret, Ander Regueiro, Francisco Campelo-Parada, Maria del Trigo, Olivier Chiche, Chekrallah Chamandi, Rishi Puri, Josep Rodes-Cabau

► To cite this version:

Vincent Auffret, Ander Regueiro, Francisco Campelo-Parada, Maria del Trigo, Olivier Chiche, et al.. Feasibility, Safety and Efficacy of Transcatheter Aortic Valve Replacement Without Balloon Pre-Dilation: A Systematic Review and Meta-Analysis. *Catheterization and Cardiovascular Interventions*, 2017, 90 (5), pp.839-850. 10.1002/ccd.27040 . hal-01636358

HAL Id: hal-01636358

<https://hal-univ-rennes1.archives-ouvertes.fr/hal-01636358>

Submitted on 16 Nov 2017

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Feasibility, Safety and Efficacy of Transcatheter Aortic Valve Replacement Without Balloon Pre-Dilation: A Systematic Review and Meta-Analysis

Vincent Auffret^{1,2}, MD, MSc, Ander Regueiro¹, MD, Francisco Campelo-Parada¹, MD, María del Trigo¹, MD, Olivier Chiche¹, MD, Chekrallah Chamandi, MD¹, Rishi Puri¹, MBBS, PhD, Josep Rodés-Cabau¹, MD.

¹Quebec Heart & Lung Institute, Laval University, Quebec City, Quebec, Canada

²Rennes 1 University, Signal and Image Processing Laboratory, Rennes, France

Short title: Feasibility, safety and efficacy of direct TAVR.

Conflict of interest statement: Dr. Josep Rodés-Cabau has received research grants from Edwards Lifesciences, Medtronic and St Jude Medical. Dr Vincent Auffret received fellowship support from the Fédération Française de Cardiologie and research grants from Abbott, Edwards Lifescience, Medtronic, Biosensors, Terumo and Boston Scientific. The other authors do not report any potential conflict of interest.

Address for correspondence:

Josep Rodés-Cabau, MD.
Quebec Heart & Lung Institute,
Laval University,
2725 Chemin Ste-Foy, G1V 4GS
Quebec City, Quebec, Canada
Phone: 418-6568711; FAX: 418-6564544
E-mail: josep.rodés@criucpq.ulaval.ca

Word count: 5036.

ABSTRACT

Objectives: To evaluate the feasibility, safety and efficacy of direct transcatheter aortic valve replacement (TAVR), i.e. TAVR without balloon pre-dilation (BPD), by performing a systematic review and meta-analysis of available evidence.

Background: Avoiding BPD during TAVR was shown to be feasible in previous studies but the risks and benefits of this technique are unknown owing to the limited number of patients included in these studies.

Methods: We performed a systematic search for studies comparing direct TAVR vs. TAVR performed with BPD. Crude risk ratios (RRs) or mean differences and 95% confidence intervals (CI) for each endpoint were calculated using random effects models.

Results: Twenty studies including 3586 patients (1606 undergoing direct TAVR) were selected for the analysis. Mean device success with direct TAVR was 88% with <5% of bail-out techniques. There were no differences between direct and BPD-TAVR in short-term (in-hospital or 30-day) mortality (RR:1.06; 95% CI:0.78-1.43) or cerebrovascular events (RR:0.92; 95% CI:0.58-1.46). Direct TAVR associated with reduced moderate or severe paravalvular leak post-TAVR (RR:0.59; 95% CI:0.36-0.98) but not with a reduced risk of permanent pacemaker implantation (RR:0.85, 95% CI:0.71-1.02). A slight increase in post-dilation was observed in direct transfemoral-TAVR recipients (RR:1.2; 95%CI:1.00-1.44).

Conclusion: Direct TAVR is feasible and safe. However, given the unadjusted nature of our results, uncertainties remain regarding the independent effect of direct TAVR on outcomes post-TAVR. Randomized studies are warranted to determine the potential benefits of direct TAVR.

Keywords: Aortic Stenosis; Transcatheter Aortic Valve Implantation; Balloon Valvuloplasty.

INTRODUCTION

Avoidance of balloon pre-dilation (BPD) during transcatheter aortic valve replacement (TAVR), a strategy known as direct TAVR, has recently emerged as part of a general shift towards simplified and more straightforward procedures¹. Although BPD was deemed mandatory as a preparatory step in the early days of TAVR, some small feasibility studies showed high procedural success rates when omitting BPD^{1, 2}. As shown in recent series³, balloon aortic valvuloplasty *per se* carries risks of permanent pacemaker implantation (PPI), severe aortic regurgitation, and stroke ranging from 0.5 to 2%, that represent the rationale supporting direct TAVR. An aggressive pre-dilation of the aortic valve and the adjacent left ventricular outflow tract may indeed increase the rate of PPI⁴ and significant paravalvular leaks (PVL) post-TAVR especially with self-expanding valves (SEV)^{5, 6}. Similarly, beyond a seemingly obvious time saving, it has been hypothesized that by reducing manipulation of the aortic arch and degenerated valve and avoiding rapid pacing runs, direct TAVR might lead to a reduction of acute cerebrovascular events (CVE) related to debris embolization or sustained impaired hemodynamic states. However, currently, data supporting direct TAVR are mainly limited to small single-center studies. Moreover, it has recently been demonstrated in a study using diffusion-weighted magnetic resonance imaging post-TAVR that direct TAVR might actually result in a higher volume of cerebral ischemic lesions⁷.

Given the current lack of consensus, we assessed the feasibility, safety and efficacy of direct TAVR by performing a systematic review and meta-analysis of available evidence.

METHODS

Search strategy

A systematic review of published literature on direct TAVR was conducted in accordance with the guidance and reporting items specified in the Preferred Reported Items for Systematic Reviews and Meta-Analysis (PRISMA) statement⁸. A computerized search was performed to identify all relevant studies from PudMed and EMBASE databases. The following keywords or terms were used: transcatheter aortic valve, TAVI, TAVR, transcutaneous aortic valve, percutaneous aortic valve, direct, predilatation, predilatation, dilation, dilatation, balloon, and valvuloplasty. The MeSH terms: Transcatheter Aortic Valve Replacement and Balloon Valvuloplasty were also used. The search strategy is outlined in the **supplemental appendix**. Databases were last accessed on July 11th 2016. In addition to the computerized search, we manually reviewed the bibliography of all included articles (V.A) to ensure complete inclusion of all possible studies.

Eligibility criteria and study selection

We deemed eligible any study of original design that compared results of direct TAVR with those of TAVR performed with BPD. We included studies in which quantitative raw data that enabled the calculation of crude risk ratios (RR) for dichotomous endpoints and mean difference (MD) for continuous endpoints were available. When potential overlapping study populations were detected (based on participating institutions and inclusion periods), the most recent publication or the publication with the most information of interest was included in the analysis. Case reports or studies published in a non-English language were excluded.

Two investigators (V.A. and A.R.) independently conducted the literature searches, study eligibility assessment and data extraction in duplicate. Any discrepancies were resolved by consensus by a third investigator (J.R.C.)

Data extraction

We extracted data of the patients and studies using a standardized data abstraction sheet. The following study-, patient- and procedure-related data were extracted from the main paper and accompanying supplemental appendix: study design; number of participating centers, region and period of enrolment, number of patients, exclusion criteria, periprocedural events definition, age, sex, baseline procedural risk assessment (by logistic European System for Cardiac Operative Risk Evaluation (EuroSCORE) or Society of Thoracic Surgeons Predicted Risk of Mortality score), number of patients with prior stroke, coronary artery disease and atrial fibrillation at baseline, access site, BEV and SEV rates, rates of patients requiring bail-out manoeuvres to cross the aortic valve or acute valve-in-valve.

Endpoints

Primary outcomes of the meta-analysis included short-term, either at discharge or at 30 days of follow-up as reported by the authors, all-cause mortality and CVE (stroke or transient ischemic attack). Secondary endpoints were rates of device success as defined by the Valve Academic Research Consortium⁹, need for balloon post-dilation, PPI, acute kidney injury (AKI) stage 2 or 3, moderate or severe PVL at discharge, and fluoroscopy time, total procedure time, total contrast used, and mean transvalvular gradient at discharge.

Statistical analysis

Crude RR and mean differences were the principal summary measures. Means and standard deviations of continuous endpoints were extracted from studies and used for the analysis. RRs were retrieved or calculated with the corresponding 95% confidence interval for each endpoint and entered in the primary analysis. Data across studies were combined using DerSimonian and Laird random effects models¹⁰. Consistency across studies was assessed with the Q-statistic and I^2 index which takes values between 0% and 100%, with values of 25% typically suggesting low, 50% moderate, and 75% large heterogeneity¹¹. Significant heterogeneity was considered present for p-values <0.10 or an I^2 >50%. Main results were

confirmed using Mantel-Haenszel fixed effect models in case of low heterogeneity. To assess the potential effect of publication bias, we inspected funnel plots for asymmetry and used the Harbord test for dichotomous outcomes and the Egger test for continuous endpoints as formal statistical tests¹². In case of significant publication bias, we intended to adjust the pooled effect estimate using the non-parametric “trim and fill” method which estimates the number and results of potential missing studies resulting from publication bias¹³. Stratified analyses were performed according to the type of implanted valve and to the timing of publication (i.e. before or after January 1st 2015) whereas sensitivity analysis was performed by including only studies that reported outcomes separately for transfemoral (TF) implantation or had <10% cases with alternative access routes in order to account for the different risk profile of patients undergoing TAVR via an alternative approach. Descriptive characteristics are presented as mean $\pm$ standard deviation for continuous variables and frequencies and percentages for categorical variables. All reported p values are 2-sided and a value of < 0.05 was considered significant. Statistical analyses were performed in STATA software (version 13.0, STATA Corp., College Station, Texas) and RevMan (Version 5.3.5, Copenhagen: The Nordic Cochrane Centre, The Cochrane Collaboration).

RESULTS

Selection of studies and population

Figure 1 shows the PRISMA flow diagram. Following removal of duplicates, 2259 records were screened at the title and abstract level. Of these, 26 studies were retrieved in full text and examined for eligibility. In addition, 29 published abstracts were also retrieved through EMBASE database and evaluated. Finally, 17 published studies^{2, 5-7, 14-26} and 3 abstracts²⁷⁻²⁹ fulfilled the inclusion criteria and were deemed eligible for the analysis (**Table 1**). Sixteen studies were observational, 4 used propensity-score matching to adjust for

imbalance in baseline characteristics between groups and 1 study was a case-control study. All but 3 studies were single-center works.

A total of 3586 patients were included in this analysis, with 1606 patients who underwent direct TAVR. Almost all studies reported the use of computed tomography during the pre-TAVR screening but only 10 studies specifically reported the imaging method used for valve sizing of which 8 used computed tomography. Balloon-expandable valves were implanted in 2274 patients without obvious differences between groups in devices iterations used in 10 out of 12 studies which reported it (**supplementary table 1**). A TF approach was used in 84.6% of cases (2959/3499 patients with known access site). The mean device success for direct TAVR was 88%. Ten studies reported rates ranging from 0.0 % to 3.8% of bail-out techniques in case of difficulty crossing the aortic valve namely rescue BPD, “buddy-balloon” technique³⁰, and partial inflation of the distal tip of the delivery balloon. Among these studies, the maximum rate of rescue BPD was 1.8%. Clinical and procedural characteristics of patients from selected studies are depicted in **Table 2**.

Impact of direct TAVR on mortality and stroke

Based on data from 16 studies (n=3260), direct TAVR was not associated with short-term all-cause mortality post-TAVR (RR: 1.06; 95% CI: 0.78-1.43, p=0.72, I²=0%) (**Figure 2, supplementary figure 1**). Similarly, no association was shown with CVE (RR: 0.92; 95% CI: 0.58-1.46, p=0.72, 14 studies, n=3143) with low statistical heterogeneity across studies (I²=4%). These results were confirmed with the use of fixed-effect models. There was no evidence of significant publication bias in the pooled estimates of primary outcomes either by funnel plot inspection or with the use of the Harbord test.

In stratified analyses by type of implanted valve, direct TAVR was associated with a significant reduction in short-term CVE in BEV recipients (RR: 0.51; 95%CI: 0.26-0.99, p=0.05, 9 studies, n=1334) but not in SEV recipients in limited data (2 studies, n=282)

(**Figures 3 and 4**). Restricting the analysis to TF TAVR recipients or by timing of publication yielded results consistent with the main analysis (**Figure 5, supplementary figure 2**).

Procedural outcomes

Fluoroscopy time was not significantly different between groups (MD: -1.91 min; 95% CI: -4.25, 0.43; $p=0.11$, 6 studies, $n=1289$), but procedural time was shorter (MD: -19.75 min; 95% CI: -36.87, -2.63; $p=0.02$, 5 studies, $n=508$) and total contrast used lower (MD: -20.77 ml; 95% CI: -28.95, -12.58; $p<0.001$, 8 studies, $n=1594$) in direct TAVR recipients. Moderate to high statistical heterogeneity was demonstrated for these endpoints. Publication biases were also shown either by funnel plot inspection or by the Egger test for all endpoints. When applying the “trim and fill” method to adjust pooled estimates, results remained unchanged for total contrast used, were consistent with the main analysis for procedure time (MD= -29.39 min; 95% CI: -46.90, -11.88; $p=0.001$) and suggested a reduced fluoroscopy time with direct TAVR (MD: -4.29 min; 95% CI: -6.96, -3.16; $p=0.002$). No data were available from SEV recipients for stratified analysis and overall data were limited in the other stratified or sensitivity analyses (**Table 3**). Amongst direct TAVR patients, reduction of volume of contrast was consistent in all analyses; fluoroscopy time was significantly reduced in BEV recipients whereas procedure time was only reduced in TF TAVR recipients and in studies published since 2015.

There was a significant association between direct TAVR and an increased rate of device success in the main analysis (RR: 1.03; 95% CI: 1.00-1.05, $p=0.02$, 13 studies, $n=2524$) without evidence of statistical heterogeneity ($I^2=5\%$) (**Figure 2**). Conversely, there was strong evidence of publication bias (p for Harbord test=0.007). When adjusting the pooled estimates using the “trim and fill” method, direct TAVR remained associated with a trend towards increased device success (RR: 1.03, 95% CI: 0.99-1.06, $p=0.06$) which was

consistent with the aggregate RR derived from a fixed effect model (RR: 1.03 95% CI: 1.00-1.06, $p=0.07$, $I^2=5\%$). Stratified and sensitivity analyses demonstrated that this relationship was mainly driven by BEV recipients and studies published since 2015 (**Figures 3 and 4, supplementary figure 2**).

Direct TAVR was not associated with the need for balloon post-dilation in all analyses performed at the exception of a slight but significant increase in TF-TAVR recipients (**Figures 2, 3 and 4, supplementary figures 1 and 2**). There was no significant statistical heterogeneity across studies or evidence of publication bias for this endpoint.

Post-procedural outcomes

The risk of AKI was not significantly different between groups in all analyses performed (**Figures 2, 3, and 4, supplementary figures 1 and 2**) with a pooled estimate in the main analysis of 1.05 (95% CI: 0.55-2.03, $p=0.88$, $I^2=28\%$, 7 studies, $n=1713$). The Harbord test suggested publication bias ($p=0.015$). However results of the “trim and fill” method were consistent with the main analysis (RR= 0.75; 95% CI: 0.38-1.48, $p=0.411$).

Based on data from 15 studies ($n=2853$), there was a trend towards a lower risk of PPI with direct TAVR (RR= 0.85, 95% CI: 0.71-1.02, $p=0.08$) (**Figure 2**). Statistical heterogeneity was low ($I^2=0\%$) and this result almost reached statistical significance when pooled estimate was derived from a fixed-effect model (RR: 0.84; 95% CI: 0.70-1.00, $p=0.05$). No publication bias was demonstrated.

One study reported the rate of moderate or severe PVL but did not report the method used for its evaluation⁷. PVL was evaluated angiographically in one study⁵ and by echocardiography in all other studies included in the analysis of this endpoint (**supplementary Figure 1**). Data synthesis of 16 studies ($n=3187$) suggested a reduced risk of moderate or severe PVL with direct TAVR (RR: 0.59; 95% CI: 0.36-0.98, $p=0.04$). Statistical heterogeneity across studies was moderate ($I^2=53\%$) but significant ($p=0.006$) and no

publication bias was demonstrated ($p=0.11$ for Harbord test). This reduction was consistent with both type of valve (**Figure 3 and 4**) and in studies published before 2015 but did not reach statistical significance in studies published since 2015 (**supplementary Figure 2**). The sensitivity analysis in TF TAVR recipients (8 studies, $n=2096$) did not show any association between direct TAVR and the rate of moderate or severe PVL (**Figure 5**).

Direct TAVR associated with reduced mean transvalvular gradient at discharge in data synthesis of 11 studies (MD: -0.54 mmHg; 95% CI: -0.95 , -0.13 , $p=0.23$, $n=2219$) (**Table 3**). There was no heterogeneity across studies ($I^2=1\%$) but publication bias was demonstrated (p for Egger test $=0.047$). Nonetheless, the pooled estimate adjusted using the “trim and fill” method was consistent with the main analysis (MD= -0.767 ; 95% CI: -1.295 ; -0.239 , $p=0.004$). No analysis suggested an increased post-procedural mean gradient in direct TAVR recipients. On the contrary, there was a trend towards a decreased gradient with direct TAVR in BEV recipients (**Table 3**).

DISCUSSION

The results of the present meta-analysis demonstrate that direct TAVR is feasible with a high rate of device success, an infrequent need for bail-out techniques, and associates with reduced procedural times and no deleterious impact on short-term mortality or stroke.

Potential risks and benefits of BPD during TAVR procedures are summarized in **Table 4**. One of the main concerns of direct TAVR is the possible difficulty in crossing the severely stenotic native aortic valve with the transcatheter valve system. The present study showed that this occurred in a very low number of patients, with $<2\%$ of cases requiring rescue BPD. Future studies with a large numbers of patients will need to determine the factors associated with direct TAVR failure. Direct TAVR associated with reduced procedural times. This may be explained by the avoidance of one procedural step (BPD), but other factors such

as facilitated valve positioning may also contribute to this reduced time. BPD is usually associated with larger aortic orificial areas in addition to increasing AR severity. This may result in more pronounced movements of the valve system during the cardiac cycle making the final positioning more challenging. However, avoiding BPD in case of important or bulky calcifications may also result in valve migration following deployment of the prosthesis. Therefore, patients' selection for direct TAVR may be of paramount importance regarding the results of this technique.

Direct TAVR associated with higher rates of device success, reflecting the lower rate of significant PVL observed with this technique in studies included in the present analysis. Whereas improved valve positioning may have played a role in such results, these associations need to be interpreted with caution due to the unadjusted nature of data analyzed in the present study. Although beneficial effects of direct TAVR *per se* cannot be ruled out, several confounders such as device improvements or growing operator experience may have influenced our results. Moreover, the non-randomized nature of included studies may have introduced bias as the selection of a direct TAVR strategy was left at the operators discretion based on pre-TAVR evaluation. As demonstrated by Abramowitz et al.¹⁴, it is therefore likely that patients with favorable anatomical characteristics of the aortic valve and arch coupled with a lower calcific burden more often underwent direct TAVR. This is of particular importance as a high correlation between the volume of calcification and the severity of PVL has been previously demonstrated³¹. Importantly, direct TAVR resulted in no deleterious effect on valve hemodynamic and no increase in the need for balloon post-dilation in the main analysis, further suggesting that current transcatheter valve systems are able to successfully open the calcified valve leaflets without the need for preparing the stenotic valve with BPD.

Transcatheter Doppler studies have shown that any mechanical interaction with a severely calcified and stenotic aortic valve, including BPD, is associated with an increase in

cerebral emboli³². It was therefore hypothesized that direct TAVR might reduce CVE by avoiding risks of debris embolization inherent with BPD³. However, a recent study challenged this hypothesis showing a higher volume of cerebral ischemic lesion on diffusion-weighted magnetic resonance imaging in patients receiving direct TAVR⁷. In the present study direct TAVR associated with similar rates of CVE in the main analysis and with a reduction of CVE in BEV recipients. Whilst these clinical results are reassuring, it could also reflect that the potential benefits from avoiding BPD may be partially thwarted by a higher mechanical interaction and subsequent embolic burden at the time of crossing the native aortic valve. Further studies should focus on the potential effects of direct TAVR on neurological events especially according to the type of valve used.

PPI occurs in 10-17% of patients in recent series using both valve types^{33, 34} and thus remains a concern given the current shift towards treating lower surgical-risk patients. Nonetheless, the valve implantation is directly responsible for less than half of new conduction disturbances during TAVR as most of them occur during BPD especially if the balloon is larger than the minor axis of the aortic annulus^{14, 35}. This has been described by Lange et al.⁴ as a two-hit model where the first hit is inflicted by a large valvuloplasty balloon to the conduction system promoting the persistence of high-degree atrioventricular block followed by a second hit by the valve frame. In their large series, Bernardi et al.²⁹ demonstrated that BPD was associated with a 1.8-fold higher risk of new-onset persistent left bundle branch block compared with direct TAVR. These data, along with the trend towards a lower risk of PPI demonstrated in the present study, suggest that direct TAVR has the potential to reduce conduction disturbances and ultimately PPI in selected high-risk patients such as in case of pre-existent right bundle branch-block or in SEV recipients.

Limitations

Several limitations of the present analysis warrant consideration. First, as previously discussed, some limitations are inherent to the comparison of BPD with direct TAVR, a technique of more recent emergence which in a non-randomized setting is inevitably confounded by patient selection, device iterations and operator experience. Moreover, most of the included studies were small single-center ones without adjusted analyses regarding the outcomes of interest. We used crude RR/MD as our principal summary measures, thereby making it impossible to identify the independent influence of direct TAVR on our endpoints. Therefore, all associations identified in the present study should be regarded as hypothesis-generating and do not allow any conclusion regarding causality. Moreover, benefits of BPD and impact of its avoidance may depend on the implanted valve type as suggested by the differing results of stratified analysis regarding CVE and PVL in the present study. These issues are currently evaluated in dedicated randomized trials (The preDilatation in tRanscathEter aortiC Valve implanTation Trial [DIRECT], NCT02448927; TAVI Without Balloon Predilatation of the Aortic Valve SAPIEN 3 [DIRECTAVI], NCT02729519). Finally, very few data exist regarding moderate BPD^{4, 14} which could represent a strategy harboring the theoretical advantages of BPD whilst limiting its risks.

CONCLUSION

Direct TAVR is a feasible and safe technique that exhibits similar rates of short-term mortality and CVE compared with TAVR with prior BPD. Future works should identify the most suitable patients for this technique to further improve its results. However, owing to inherent limitations of studies included in this meta-analysis, the independent effect of direct TAVR remains uncertain and a randomized evaluation comparing this technique to conventional BPD-TAVR is warranted.

REFERENCES

1. Grube E, Naber C, Abizaid A, Sousa E, Mendiz O, Lemos P, Kalil Filho R, Mangione J, Buellesfeld L. Feasibility of transcatheter aortic valve implantation without balloon predilation: A pilot study. *JACC Cardiovasc Interv* 2011;4:751-757
2. Mollmann H, Kim WK, Kempfert J, Blumenstein J, Liebetrau C, Nef H, Van Linden A, Walther T, Hamm C. Transfemoral aortic valve implantation of edwards sapien xt without predilatation is feasible. *Clin Cardiol* 2014;37:667-671
3. Eltchaninoff H, Durand E, Borz B, Furuta A, Bejar K, Canville A, Farhat A, Fraccaro C, Godin M, Tron C, Sakhuja R, Cribier A. Balloon aortic valvuloplasty in the era of transcatheter aortic valve replacement: Acute and long-term outcomes. *Am Heart J* 2014;167:235-240
4. Lange P, Greif M, Vogel A, Thaumann A, Helbig S, Schwarz F, Schmitz C, Becker C, D'Anastasi M, Boekstegers P, Pohl T, Laubender RP, Steinbeck G, Kupatt C. Reduction of pacemaker implantation rates after corevalve(r) implantation by moderate predilatation. *EuroIntervention* 2014;9:1151-1157
5. Fiorina C, Maffeo D, Curello S, Lipartiti F, Chizzola G, D'Aloia A, Adamo M, Mastropierro R, Gavazzi E, Ciccarese C, Chiari E, Ettori F. Direct transcatheter aortic valve implantation with self-expandable bioprosthesis: Feasibility and safety. *Cardiovasc Revasc Med* 2014;15:200-203
6. Toutouzas K, Latsios G, Stathogiannis K, Drakopoulou M, Synetos A, Sanidas E, Mastrokostopoulos A, Trantalís G, Kaitozis O, Lazaros G, Yucel S, Gerckens U, Grube E, Tousoulis D. One-year outcomes after direct transcatheter aortic valve implantation with a self-expanding bioprosthesis. A two-center international experience. *Int J Cardiol* 2016;202:631-635
7. Bijuklic K, Haselbach T, Witt J, Krause K, Hansen L, Gehrckens R, Rieß FC, Schofer J. Increased risk of cerebral embolization after implantation of a balloon-expandable aortic valve without prior balloon valvuloplasty. *JACC Cardiovasc Interv* 2015;8:1608-1613
8. Moher D, Liberati A, Tetzlaff J, Altman DG. Preferred reporting items for systematic reviews and meta-analyses: The prisma statement. *Open Med* 2009;3:e123-130

9. Kappetein AP, Head SJ, Gnreux P, Piazza N, Van Mieghem NM, Blackstone EH, Brott TG, Cohen DJ, Cutlip DE, Van Es GA, Hahn RT, Kirtane AJ, Krucoff MW, Kodali S, Mac KMJ, Mehran R, Rods-Cabau J, Vranckx P, Webb JG, Windecker S, Serruys PW, Leon MB. Updated standardized endpoint definitions for transcatheter aortic valve implantation: The valve academic research consortium-2 consensus document. *Eur Heart J* 2012;33:2403-2418b
10. DerSimonian R, Laird N. Meta-analysis in clinical trials. *Control Clin Trials*. 1986;7:177-188
11. Higgins JP, Thompson SG. Quantifying heterogeneity in a meta-analysis. *Stat Med* 2002;21:1539-1558
12. Harbord RM, Egger M, Sterne JA. A modified test for small-study effects in meta-analyses of controlled trials with binary endpoints. *Stat Med* 2006;25:3443-3457
13. Duval S, Tweedie R. Trim and fill: A simple funnel-plot-based method of testing and adjusting for publication bias in meta-analysis. *Biometrics*. 2000;56:455-463
14. Abramowitz Y, Jilaihawi H, Chakravarty T, Kashif M, Matar G, Hariri B, Patel J, Sharma RP, Cheng W, Makkar RR. Feasibility and safety of balloon-expandable transcatheter aortic valve implantation with moderate or without predilatation. *EuroIntervention* 2016;11:1132-1139
15. Conradi L, Schaefer A, Seiffert M, Schirmer J, Schaefer U, Schon G, Blankenberg S, Reichenspurner H, Treede H, Diemert P. Transfemoral tavi without pre-dilatation using balloon-expandable devices: A case-matched analysis. *Clin Res Cardiol* 2015;104:735-42.
16. Conradi L, Seiffert M, Schirmer J, Koschyk D, Blankenberg S, Reichenspurner H, Diemert P, Treede H. Transapical transcatheter aortic valve implantation without prior balloon aortic valvuloplasty: Feasible and safe. *Eur J Cardiothorac Surg* 2014;46:61-66
17. Islas F, Almeria C, Garcia-Fernandez E, Jimenez P, Nombela-Franco L, Olmos C, Marcos-Alberca P, Cuadrado A, Fernandez-Ortiz A, Macaya C, Perez de Isla L. Usefulness of echocardiographic criteria for transcatheter aortic valve implantation without balloon predilatation: A single-center experience. *J Am Soc Echocardiogr* 2015;28:423-429
18. Kempfert J, Meyer A, Kim WK, van Linden A, Arsalan M, Blumenstein J, Mollmann H, Walther T. First experience without pre-balloon in transapical aortic valve implantation: A propensity score-matched analysis. *Eur J Cardiothorac Surg* 2015;47:31-38

19. Kim WK, Praz F, Blumenstein J, Liebetrau C, Gaede L, Van Linden A, Hamm C, Walther T, Windecker S, Möllmann H. Transfemoral aortic valve implantation of edwards sapien 3 without predilatation. *Catheter Cardiovasc Interv* 2016 doi: 10.1002/ccd.26464.
20. Kochman J, Kołtowski Ł, Huczek Z, Scisło P, Bakoń L, Wilimski R, Rymuza B, Opolski G. Direct transcatheter aortic valve implantation - one-year outcome of a case control study. *Postepy Kardiologii Interwencyjnej* 2014;10:250-257
21. Wong SC, Pawar S, Minutello RM, Horn EM, Skubas NJ, Devereux RB, Salemi A. Device success and 30-day clinical outcome in patients undergoing preimplant valvuloplasty in transfemoral versus omitting valvuloplasty in transapical transcatheter aortic valve replacement. *J Thorac Cardiovasc Surg* 2015;150:1111-1117
22. Van Linden A, Kim WK, Liebetrau C, Doss M, Hecker F, Arsalan M, Blumenstein J, Möllmann H, Walther T. Simplified transapical aortic valve implantation using the sapien 3 valve without preballooning. *Innovations (Phila)* 2015;10:406-409
23. Wendler O, Dworakowski R, Monaghan M, MacCarthy PA. Direct transapical aortic valve implantation: A modified transcatheter approach avoiding balloon predilatation. *Eur J Cardiothorac Surg* 2012;42:734-736
24. Aggarwal SK, Delahunty N, Wong B, Tamimi AN, Reinthaler M, Cheang MH, Roberts N, Yap J, Ozkor M, Mullen MJ. Balloon-expandable transcatheter aortic valves can be successfully and safely implanted transfemorally without balloon valvuloplasty. *J Interv Cardiol* 2016;29:319-324
25. Bandali A, Parry-Williams G, Kassam A, Palmer S, Williams PD, de Belder MA, Owens A, Goodwin A, Muir DF. Direct transfemoral transcatheter aortic valve implantation without balloon pre-dilatation using the edwards sapien xt valve. *Catheter Cardiovasc Interv* 2016 doi: 10.1002/ccd.26564
26. Pagnesi M, Jabbour RJ, Latib A, Kawamoto H, Tanaka A, Regazzoli D, Mangieri A, Montalto C, Ancona MB, Giannini F, Chieffo A, Montorfano M, Monaco F, Castiglioni A, Alfieri O, Colombo A. Usefulness of predilation before transcatheter aortic valve implantation. *The Am J Cardiol* 2016;118:107-112

27. Lettieri C, Romano M, Baccaglioni N, Buffoli F, Rosiello R, Agostini F, Fiorani V, Pacchioni R, Rambaldini M, Zanini R. Transcatheter aortic valve implantation without balloon predilation: A non-randomized single centre experience. *G Ital Cardiol* 2014;15:e30
28. Nielsen NE, Ahn H, Baranowski J, Dahlin LG, Nylander E, Wallby L. Transcatheter aortic valve implantation without prior balloon dilatation-a non-randomized single centre experience. *J Am Coll Cardiol* 2013;62:B218
29. Bernardi FL, Ribeiro HB, Carvalho LA, Sarmento-Leite R, Mangione JA, Lemos PA, Abizaid A, Grube E, Rodes Cabau J, De Brito Jr FS. Direct transcatheter heart valve implantation versus implantation with balloon predilatation: Insights from the brazilian tavr registry. *J Am Coll Cardiol* 2016;67:415
30. Noble S, Roffi M. Retrograde aortic valve crossing of the corevalve prosthesis using the buddy balloon technique. *Catheter Cardiovasc Interv* 2014;84:897-899
31. Khaliq OK, Hahn RT, Gada H, Nazif TM, Vahl TP, George I, Kalesan B, Forster M, Williams MB, Leon MB, Einstein AJ, Pulerwitz TC, Pearson GD, Kodali SK. Quantity and location of aortic valve complex calcification predicts severity and location of paravalvular regurgitation and frequency of post-dilation after balloon-expandable transcatheter aortic valve replacement. *JACC Cardiovasc Interv* 2014;7:885-894
32. Kahlert P, Al-Rashid F, Dottger P, Mori K, Plicht B, Wendt D, Bergmann L, Kottenberg E, Schlamann M, Mummel P, Holle D, Thielmann M, Jakob HG, Konorza T, Heusch G, Erbel R, Eggebrecht H. Cerebral embolization during transcatheter aortic valve implantation: A transcranial doppler study. *Circulation* 2012;126:1245-1255
33. Holmes DR, Jr., Nishimura RA, Grover FL, Brindis RG, Carroll JD, Edwards FH, Peterson ED, Rumsfeld JS, Shahian DM, Thourani VH, Tuzcu EM, Vemulapalli S, Hewitt K, Michaels J, Fitzgerald S, Mack MJ. Annual outcomes with transcatheter valve therapy: From the sts/acc tvf registry. *J Am Coll Cardiol* 2015;66:2813-2823
34. Walther T, Hamm CW, Schuler G, Berkowitsch A, Kotting J, Mangner N, Mudra H, Beckmann A, Cremer J, Welz A, Lange R, Kuck KH, Mohr FW, Mollmann H, board Ge.

- Peri-operative results and complications in 15,964 transcatheter aortic valve implantations from the german aortic valve registry (gary). *J Am Coll Cardiol* 2015;65:2173-80
35. Nuis RJ, Van Mieghem NM, Schultz CJ, Tzikas A, Van Der Boon RM, Maugeness AM, Cheng J, Piazza N, Van Domburg RT, Serruys PW, De Jaegere PP. Timing and potential mechanisms of new conduction abnormalities during the implantation of the medtronic corevalve system in patients with aortic stenosis. *Eur Heart J* 2011;32:2067-2074

FIGURES LEGENDS

Figure 1- Flow chart of selected studies

Flow diagram –based on the PRISMA statement- of included studies.

Figure 2- Summary RRs of dichotomous endpoints (main analysis).

Forest plot of summary RRs for the comparison of direct TAVR versus TAVR performed with BPD. RRs lower than 1 favor direct TAVR except for device success for which a RR greater than 1 favor direct TAVR. AKI=Acute Kidney Injury; BPD=Balloon pre-dilation; CVE=Cerebrovascular events; PPI=Permanent pacemaker implantation; PVL=Paravalvular leak; TAVR=Transcatheter aortic valve replacement.

Figure 3- Summary RRs of dichotomous endpoints (stratified analysis-BEV recipients)

Forest plot of summary RRs for the comparison of direct TAVR versus TAVR performed with BPD in BEV recipients. Interpretation as in Figure 2. BEV=Balloon-expandable valve. Other abbreviations as in Figure 2.

Figure 4- Summary RRs of dichotomous endpoints (stratified analysis-SEV recipients)

Forest plot of summary RRs for the comparison of direct TAVR versus TAVR performed with BPD in SEV recipients. Interpretation as in Figure 2. SEV=Self-expandable valve. Other abbreviations as in Figure 2.

Figure 5- Summary RRs of dichotomous endpoints (sensitivity analysis-TF TAVR)

Forest plot of summary RRs for the comparison of direct TAVR versus TAVR performed with BPD in TF-TAVR recipients. Interpretation as in Figure 2. TF=Transfemoral. Other abbreviations as in Figure 2.

Table 1 - Characteristics of selected studies.

Author ^{Ref}	Year	Region	Centers (n)	Sample size (n)	Design	Inclusion period	Exclusion criteria / Direct TAVR criteria	Periprocedural events criteria
Abramowitz ¹⁵	2016	USA	1	513	Observational	April 2012-December 2014	Valve-in-Valve, Prior BAV within 30d, Regular BPD*.	VARC-2
Aggarwal ²⁵	2016	UK	1	154	Observational	March 2012-July 2014	Concomitant intervention, Valve-in-Valve.	VARC-2
Bandali ²⁶	2016	UK	1	81	Observational	November 2010-March 2013	Valve-in-Valve. Excluded from direct TAVR: BPD deemed necessary for assessment of potential coronary obstruction or valve sizing, extreme calcification.	VARC
Bernardi ³⁰	2016	Brazil	22	761	Observational	January 2008-January 2015	Valve-in-Valve, Use of Innovare valve or ESV XT via TA approach.	VARC-2
Bijuklic ⁸	2015	Germany	1	87	Observational	ND	Refusal of or contra-indications to DW-MRI, CVE within the prior 6 months, renal failure, presentation with cardiogenic shock or severe hypotension, need for any other cardiac surgical or interventional procedure during or after TAVR procedure before DW-MRI, clinical apparent stroke within 3 days post-TAVR.	VARC-2

Conradi ¹⁶	2015	Germany	1	52	PS matching	ND	Non-TF TAVR.	VARC-2
Conradi ¹⁷	2014	Germany	1	100	Observational	May 2011-December 2012	TF-TAVR.	VARC-2
Fiorina ⁶	2014	Italy	1	100	Observational	June 2011-June 2013	Pure aortic regurgitation, Valve-in-Valve, bicuspid aortic valve, prior BAV, BPD. performed to exclude coronary occlusion.	VARC-2
Isas ¹⁸	2015	Spain	1	249	Observational	January 2009-August 2014	Direct TAVR choice based on fulfillment of institutional echocardiographic criteria.	VARC-2
Kempfert ¹⁹	2015	Germany	1	80	PS matching	March 2012-July 2013	ND	ND
Kim ²⁰	2016	Germany, Switzerland	2	163	Observational	March 2014-July 2015	Unsuitable anatomy for ESV S3 prosthesis.	VARC-2
Kochman ²¹	2014	Poland	1	24	Case-control study	March 2010-April 2013	No control case found.	VARC-2
Lettieri ²⁸	2014	Italy	1	72	Observational	ND	Routine direct TAVR if severe left ventricular dysfunction.	ND
Mollman ³	2014	Germany	1	56	Observational	ND	Exclusion based on European Society of Cardiology guidelines; annular plane to coronary ostia < 8mm; annulus diameter > 27 mm.	VARC-2
Nielsen ²⁹	2013	Sweden	1	160	Observational	September 2008-April 2013	ND	ND
Pagnesi ²⁷	2016	Italy	1	517	PS matching	November 2007-October 2015	Non-TF TAVR, patients not receiving a SAPIEN XT/SAPIEN 3 valve or a CoreValve/Evolut R	VARC-2

Toutouzas ⁷	2016	Greece, Germany	2	210	Observational	January 2008-September 2013	ND	VARC-2
Van Linden ²³	2015	Germany	1	66	Observational	2013-2014	ND	ND
Wendler ²⁴	2012	UK	1	20	Observational	ND	ND	VARC
Wong ²²	2015	USA	1	121	Observational	May 2012-December 2013	Valve-in-valve; patients included in a protocol that mandated BPD.	VARC-2

*BPD group consisted of “moderate” BPD with a mean pre-dilation balloon diameter/CT mean annulus diameter of 0.65/1.

BAV=Balloon aortic valvuloplasty; BPD= Balloon pre-dilation; CVE=Cerebrovascular events; DW-MRI= Diffusion-weighted magnetic resonance imaging; ESV S3= Edwards valve Sapien 3; ESV XT= Edwards valve Sapien XT; ND= No Data; PS=Propensity score; TA= Transapical; TAVR=Transcatheter aortic valve replacement; TF=Transfemoral; UK= United Kingdom; USA= United States of America; VARC= Valve Academic Research Consortium.

Table 2- Clinical and procedural characteristics of patients from selected studies

Author ^{ref}	Age (years)	Male Sex (%)	Surgical risk score	Prior stroke (%)	CAD (%)	AF (%)	Approach (%)	Valve type (%)	Bail-out maneuvers for valve crossing (%)*	TV-in-TV (%)*
Abramowitz ¹⁵	82.5±8.6	61.8	8.4±5.2	14.0	64.7	ND	TF (83); TA (6); TAO (10); SC (1)	ESV (100)	0.8 vs. 1.5	1.7 vs. 4.8
Aggarwal ²⁵	81.4	57.1	19.8 [†]	22.4	ND	ND	TF (100)	ESV (100)	3.8 vs. 0.0	ND
Bandali ²⁶	83.7	61.7	23.7 [†]	ND	53.0	30.9	TF (100)	ESV (100)	ND	ND
Bernardi ³⁰	81.8±7.1	48.6	10.2±7.9	ND	58.1	13.0	TF (97); non-TF (3)	ESV (24); MCV (76)	2.3 vs. 0.0	5.1 vs. 4.3
Bijuklic ⁸	83.2±6.2	49.4	22.2±15.4 [†]	16.1	67.8	32.1	ND	ESV (100)	ND	ND
Conradi ¹⁶	81.5±5.8	63.5	5.5±2.5	13.5	51.0	23.1	TF (100)	ESV (100)	0.0 vs. 0.0	3.8 vs. 3.8
Conradi ¹⁷	79.5±7.5	51.0	8.0±6.0	17.0	71.0	51.0	TA (100)	ESV (100)	0.0 vs. 0.0	2.0 vs. 2.0
Fiorina ⁶	83.0±7.5	47.9	8.7±6.2	ND	44.8	ND	TF (66); TAO (21); SC (13)	MCV (100)	1.8 vs. 0.0	3.6 vs. 4.0
Islas ¹⁸	82.7±5.6	35.0	18.1±9.7	ND	ND	ND	TF (100)	ESV (67); MCV (33)	ND	3.8 vs. 5.3
Kempfert ¹⁹	79.5	70.0	7.42	8.5	82.5	35.0	TA (100)	ESV (100)	ND	2.5 vs. 2.5
Kim ²⁰	81.9±3.5	53.4	4.3±1.5	ND	58.9	ND	TF (100)	ESV (100)	1.3 vs. 0.0	ND
Kochman ²¹	81.6±5.3	62.5	19.0±6.6 [†]	16.7	54.2	41.7	TF (75); SC (25)	MCV (100)	ND	ND
Lettieri ²⁸	84.0±5.0	56.0	22.9±10 [†]	ND	ND	ND	TF (90); TAO (4); SC (6)	MCV (100)	ND	ND
Mollman ³	81.9±5.9	50.0	6.0±2.9	16.1	60.7	41.1	TF (100)	ESV (100)	0.0 vs. ND	0.0 vs. ND
Nielsen ²⁹	80.0	46.9	6.2	ND	ND	ND	TF (69); TA (30); TAO (1)	ESV (100)	ND	ND
Pagnesi ²⁷	80.2±7.4	36.7	21.1±16.0	ND	41.1	31.3	TF (100)	ESV (52); MCV (48)	ND	ND
Toutouzas ⁷	80.7±10.7	39.5	19.9±15.4 [†]	ND	23.8	ND	TF (92); TAO (1); SC (7)	MCV (100)	ND	ND

Van Linden ²³	83.0±6.7	61.7	7.5±5.2	16.5	74.7	45.6	TA (100)	ESV (100)	0.0 vs. ND	0.0 vs 0.0
Wendler ²⁴	82.0±3.0	77.0	30.0±2.0 [†]	50.0	ND	ND	TA (100)	ESV (100)	ND	ND
Wong ²²	84.4±7.1	49.0	8.9±4.9	15.7	ND	39.7	TF (59); TA (41)	ESV (100)	0.0 vs. 0.0	ND

* Direct TAVR group vs. BPD group

[†] Logistic EuroSCORE I if [†]; Society of Thoracic Surgeon Predicted Risk of Mortality otherwise.

[‡] Baseline characteristics were unavailable for the matched cohorts and thus were retrieved from the entire cohort (n=206).

AF=Atrial fibrillation; BPD=Balloon pre-dilation; CAD=Coronary artery disease; ESV= Edwards Sapien Valves; MCV= Medtronic Corevalve; ND=No Data;TA=Transapical; TAo= Transaortic; TF=Transfemoral; TV-in-TV=Transcatheter valve within a transcatheter valve; SC= Sub-clavian.

Table 3-Results of analyses for continuous endpoints

Endpoint	Number of studies	Number of patients	MD (95% CI)	p-value	I-squared
Main analysis					
Fluoroscopy time, min	6	1289	-1.91 (-4.25, 0.43)	0.11	84%
Procedure time, min	5	508	-19.75 (-36.87, -2.63)	0.02	88%
Total contrast used, ml	8	1594	-20.77 (-28.95, -12.58)	<0.001	40%
Post-procedural mean gradient, mm Hg	11	2219	-0.54 (-0.95, -0.13)	0.01	1%
BEV patients					
Fluoroscopy time, min	5	772	-2.49 (-4.96, -0.02)	0.05	84%
Procedure time, min	4	259	-18.05 (-41.74, 5.64)	0.14	90%
Total contrast used, ml	6	828	-22.64 (-33.99, -11.30)	<0.001	56%
Post-procedural mean gradient, mm Hg	6	909	-0.58 (-1.24, 0.08)	0.09	0%
SEV patients					
Fluoroscopy time, min	0	0	-	-	-
Procedure time, min	0	0	-	-	-
Total contrast used, ml	0	0	-	-	-
Post-procedural mean gradient, mm Hg	2	234	0.88 (-0.45, 2.21)	0.20	0%
Studies published before January 1st 2015					
Fluoroscopy time, min	2	120	-1.05 (-3.55, -1.46)	0.41	61%
Procedure time, min	2	120	-24.96 (-68.07, 18.16)	0.26	97%
Total contrast used, ml	3	176	-26.28 (-40.75, -11.82)	<0.001	18%
Post-procedural mean gradient, mm Hg	3	180	-0.53 (-1.82, 0.76)	0.42	0%
Studies published after January 1st 2015					
Fluoroscopy time, min	4	1169	-2.28 (-5.47, 0.92)	0.16	83%
Procedure time, min	3	388	-18.12 (-28.95, -7.28)	0.001	30%
Total contrast used, ml	5	1418	-18.34 (-27.97, -8.72)	0.002	42%
Post-procedural mean gradient, mm Hg	8	2039	-0.43 (-0.97, 0.11)	0.12	23%
TF cohort					
Fluoroscopy time, min	2	569	-1.64 (-7.02, 3.75)	0.55	83%
Procedure time, min	2	301	-21.41 (-34.30, -8.52)	0.001	30%
Total contrast used, ml	4	874	-21.46 (-30.25, -12.66)	<0.001	0%
Post-procedural mean gradient, mm Hg	6	1395	-0.37 (-1.04, 0.29)	0.27	28%

BEV= Balloon –expandable valve; CI= confidence interval; MD= Mean difference between direct TAVR and TAVR with BPD; SEV= Self-expandable valve. Other abbreviations as in Table 2.

Table 4- Potential advantages and drawbacks of balloon predilation during TAVR procedures

Advantages	Drawbacks
Improved sizing of the aortic annulus	Acute aortic regurgitation before valve implantation
Identification of patients at high-risk of coronary occlusion	Conduction disturbances: left bundle branch and atrioventricular blocks
Easier crossing of the aortic valve	Cerebrovascular events
May reduce the risk of valve malposition by minimizing radial counterforces	Annulus rupture
Allow optimal expansion of the prosthesis	May increase valve movement during the positioning phase
Reduced need for postdilation: may improve prosthesis durability	Prolongs the procedure
Training of young operators	Increased need for rapid pacing