Ensuring Robust and Tissue-Independent Operation of Implantable, Ingestible, and Injectable Antennas

D. Nikolayev, W. Joseph, M. Zhadobov, Ronan Sauleau, L. Martens

► To cite this version:

D. Nikolayev, W. Joseph, M. Zhadobov, Ronan Sauleau, L. Martens. Ensuring Robust and Tissue-Independent Operation of Implantable, Ingestible, and Injectable Antennas. 13th International Symposium on Medical Information and Communication Technology, ISMICT 2019, May 2019, Oslo, Norway. pp.8743831, 10.1109/ISMICT.2019.8743831 . hal-02281669

HAL Id: hal-02281669

<https://hal-univ-rennes1.archives-ouvertes.fr/hal-02281669>

Submitted on 2 Nov 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Distributed under a Creative Commons Attribution| 4.0 International License

Ensuring Robust and Tissue-Independent Operation of Implantable, Ingestible, and Injectable Antennas

Denys Nikolayev
MAG – SCI-STI-AS
EPFL
Lausanne, Switzerland
d@deniq.com

Wout Joseph
INTEC – WAVES
imec / Ghent University
Ghent, Belgium
Wout.Joseph@UGent.be

Maxim Zhadobov
IETR – UMR CNRS 6164
Université de Rennes 1
Rennes, France
Maxim.Zhadobov@univ-rennes1.fr

Ronan Sauleau
IETR – UMR CNRS 6164
Université de Rennes 1
Rennes, France
Ronan.Sauleau@univ-rennes1.fr

Luc Martens
INTEC – WAVES
imec / Ghent University
Ghent, Belgium
Luc1.Martens@ugent.be

Abstract—In-body bioelectronics relies on antennas to interface with external on- or off-body equipment. In this paper, we address the impedance detuning of in-body antennas caused by nondeterministic electromagnetic properties of biological tissues. The impedance robustness can be improved in two ways. First, by using wideband and multiresonance antennas with an added benefit of higher achievable data rates. Second, by applying specific dielectric-loaded narrow-band microstrip designs that additionally show enhanced radiation performance and robustness to inner components. Fully tissue-independent characteristics can be realized using hybrid slot-patch designs.

Keywords—antenna, bioelectronics, implantable, in-body, ingestible, injectable, microstrip antenna, robustness

I. INTRODUCTION

Autonomous implantable, ingestible, and injectable (in-body) bioelectronics offer powerful capabilities in medicine, clinical research, occupational health, and professional sports [1]–[3]. Specifically, wireless *biotelemetry* allows for monitoring of a wide range of human or animal physiological parameters: vitals, blood tests, and so on. *Precision medicine* requires precision diagnostics [4], and miniature implants make it possible to yield more precise information than ever before about one's health (Fig. 1). Neural interfaces allow us to study the brain via mapping, assisting, augmenting, and repairing cognitive or sensory-motor functions [5]. The emerging concept of *electroceuticals* aims individual nerve fibers or specific neural circuits that regulate organs to treat an array of conditions [6].

In-body bioelectronics uses radiofrequency (RF) antennas to communicate with external on- or off-body systems. A wide range of RF antennas has been proposed for body-implantable applications [3]. So far, establishing a robust link between an

Fig. 1. Application example (biotelemetry) of implantable, ingestible, and injectable wireless bioelectronics.

in-body device and external equipment remains a major challenge because of the low total efficiencies ($\eta_{\text{tot}} < 0.1\%$) of the antennas operating in lossy media with uncertain electromagnetic (EM) properties. In addition to improved communication link, efficient and robust in-body antennas allow for minimization of measurement errors in path loss characterization *in vivo*, especially for deep-body locations in large animals.

In this paper, we 1) review the design challenges and 2) provide guidelines on how to ensure the robust and tissue-independent operation of in-body antennas.

II. DESIGN CHALLENGES

The radiation performance of in-body antenna strongly depends on dispersive EM properties of surrounding tissues [7] (characterized by the complex permittivity $\tilde{\epsilon}(\omega) = \epsilon' - j\epsilon''$, where $\omega = 2\pi f$ is the angular frequency). The evanescent fields dissipate and the radiated power attenuates due to tissue losses

This publication is supported by the European Union through the European Regional Development Fund (ERDF), and by the French region of Brittany, Ministry of Higher Education and Research, Rennes Métropole and Conseil Départemental 35, through the CPER SOPHIE/STIC & Ondes.

$\varepsilon'' = \sigma/\omega$ (where σ is the conductivity). The attenuation α is proportional to frequency f .

The propagating wave scatters because of the significant tissue heterogeneity $\varepsilon(\mathbf{r})$ and high wave-impedance contrast ΔZ between the tissues and free space; $\Delta Z \propto 1/f$ [8]. Therefore, for a given electrical size ka of the antenna, the achievable radiation efficiency η can be improved by choosing an optimal operating frequency f_{opt} . The optimal frequency is a tradeoff between α on the one hand and ΔZ on the other [9].

Caused by heterogeneous and nondeterministic nature of tissue effective-EM properties [8], the antenna impedance detuning strongly affects the *total efficiency* of the antenna $\eta_{\text{tot}} = \eta(1 - |\Gamma|^2)$, where Γ is the reflection coefficient. The conductivity σ of surrounding tissues affects the bandwidth and η [10]. The relative permittivity ε_r loads the antenna reducing its resonant frequency. Note that this dielectric loading also increases the electrical size ka of the antenna improving η [11].

The wider is the range of EM properties the antenna can handle (i.e. matching remains below $|S_{11}| \leq -10$ dB), the more it is considered *robust*. *Tissue-independent* antenna, therefore, can operate in all body tissues [12]. Such antenna allows for its application in a wide range of scenarios. For instance, the same device can be used as an ingestible for humans and as an implantable for animals that can range in size from a mouse to a cow [13].

An in-body device typically contains various electronic and biomedical circuits, sensors, and a power source. These inner components can affect the antenna impedance performance as well and must be taken into account during the numerical analysis of the antenna.

III. ENSURING ROBUST OPERATION

To analyze the impedance robustness of the antenna to tissues, we use an isotropic homogeneous $\varnothing 100$ -mm spherical phantom (Fig. 2) [14]. The EM properties (ε_r , σ) are parametrized to span the whole range of tissue EM properties: from fat (the lowest) to cerebrospinal fluid (the highest) [8].

A. Design Approaches

1) *Wideband antennas*—The most common approach to improving detuning immunity of in-body antennas is to yield a sufficiently large bandwidth. Ultra-wideband designs have been proposed recently [15]–[17]. In addition to improved robustness, such antennas allow for higher achievable data rates. The ultra-wideband designs typically ensure operation in muscle and gastrointestinal tract (GI) tissues.

Going beyond this range of EM properties requires further increasing the bandwidth. For a given ka , this usually involves increasing antenna losses by, for instance, reinforcing the coupling with surrounding tissues. Obviously, this leads to deterioration of the radiation efficiency. With rare exceptions, wideband in-body antennas are based on loop designs that could be sensitive to the inner components of the device due to the absence of the ground plane.

Fig. 2. Evaluation of impedance robustness to tissues (phantom not to scale). The antenna is centered in a $\varnothing 100$ -mm spherical phantom with varying EM properties.

Fig. 3. Reflection coefficient $|S_{11}|$ of the slot-patch antenna computed in air and in a $\varnothing 100$ -mm spherical phantom (Fig. 2) with muscle- and fat-equivalent EM properties.

2) *Multiresonance antennas*—An alternative to wideband approach would be to design an antenna with multiple resonances covering the EM properties of the tissues surrounding the antenna ranging from high to low water content. This multiresonance method was implemented in [12]. The antenna is based on folded-meandered dipole design and shows good matching ($|S_{11}| < -10$ dB) in all high-water-content tissues (note that these include all gastrointestinal tract tissues). The matching in fat is $|S_{11}| \approx -8$ dB.

3) *Dielectric-loaded microstrip antennas*—Having relatively narrow bandwidths, specific dielectric-loaded microstrip designs allow reducing the effect of uncertain EM properties on the antenna impedance. Representing a microstrip antenna as a combination of transmission line sections [18], one can see that the resonance frequency of the antenna depends on the characteristic impedance Z_0 of each microstrip section. This characteristic impedance is defined mostly by the

invariable geometry and EM properties of the substrate and superstrate [19]. The E-field fringing into the tissues with variable EM properties accounts for the antenna detuning. These fringing fields can be partially withdrawn from the tissues by using high-permittivity superstrates of sufficient thickness.

In addition to the enhanced robustness, this approach does not require sacrificing the radiation efficiency of the antenna. On the contrary, it allows improving the theoretically achievable efficiency by increasing the electrical size of the antenna. This approach was demonstrated in [18] using a miniature 434-MHz microstrip antenna (Fig. 2) loaded with a 0.5-mm-thick alumina superstrate ($\epsilon_r \approx 10$). The antenna shows good matching in all high-water-content tissues ($|S_{11}| < -10$ dB) having a radiation efficiency of 0.4% and gain $G = -22$ dBi (in a $\phi 100$ -mm muscle-equivalent phantom, Fig. 2).

4) *Hybrid slot-patch antennas*—Dielectric loading of microstrip antennas requires costly materials that limit the scope of potential applications. Using low-cost materials (usually, lower-permittivity biocompatible plastics) results in stronger antenna coupling to the surrounding lossy tissues and, therefore, reduced robustness. Sections of microstrip antennas with high characteristic impedance (high- Z_0) are the most affected because of larger fringing fields penetrating into the tissues.

Realizing the high- Z_0 sections as coplanar-slot lines in the ground plane moves the fringing fields inside of the low-loss device and away from the tissues. This helps to reduce the variance of effective EM properties in the near-field zone of the antenna hence results in significantly improved robustness at a cost of the radiation performance. Fig. 3 shows the impedance performance of a novel hybrid slot-patch antenna [13].

The proposed design remains well matched in both air and body tissues, which is, to the best of our knowledge, the first of its kind. However, the radiation performance of the antenna ($G = -34$ dBi) stays comparable with counterparts [9], but it is less efficient than the dielectric-loaded antennas using materials with $\epsilon_r \geq 10$.

IV. CONCLUSION

We reviewed the design challenges and approaches to tissue-independent wireless operation of implantable bioelectronics. Robust antenna characteristics can be achieved using 1) wideband and multiresonance antennas with an added benefit of improved data rates or 2) dielectric-loaded narrow-band microstrip designs showing better radiation performance and robustness to inner components.

Fully tissue-independent characteristics can be realized using hybrid slot-patch designs. Such antennas are promising candidates for subject-independent applications. For instance, it can be used as an ingestible for humans or as an implantable for a wide range of animals: from rodents to cattle.

Moreover, robust and efficient in-body antennas allow for minimization of measurement errors in path loss characterization *in vivo*, especially for deep-body applications or in large animals [20].

The considered antenna designs (both in planar and conformal realizations) can also find their applications in other non-medical fields where establishing robust communication links is required with devices in high-permittivity and lossy environments.

REFERENCES

- [1] A. S. Y. Poon, "Miniaturized Biomedical Implantable Devices," in *Implantable Bioelectronics*, Weinheim, Germany: Wiley-VCH, 2014, pp. 45–64.
- [2] A. Kiourti and K. S. Nikita, "A review of in-body biotelemetry devices: implantables, ingestibles, and injectables," *IEEE Trans. Biomed. Eng.*, vol. 64, no. 7, pp. 1422–1430, Jul. 2017.
- [3] D. Nikolayev, M. Zhadobov, R. Sauleau, and P. Karban, "Antennas for ingestible capsule telemetry," in *Advances in Body-Centric Wireless Communication: Applications and State-of-the-Art*, London, UK: IET, 2016, pp. 143–186.
- [4] S. J. Aronson and H. L. Rehm, "Building the foundation for genomics in precision medicine," *Nature*, vol. 526, no. 7573, pp. 336–342, Oct. 2015.
- [5] J. J. Jun *et al.*, "Fully integrated silicon probes for high-density recording of neural activity," *Nature*, vol. 551, no. 7679, pp. 232–236, Nov. 2017.
- [6] K. Famm, B. Litt, K. J. Tracey, E. S. Boyden, and M. Slaoui, "Drug discovery: A jump-start for electroceuticals," *Nature*, vol. 496, pp. 159–161, Apr. 2013.
- [7] D. Nikolayev, M. Zhadobov, P. Karban, and R. Sauleau, "Electromagnetic radiation efficiency of body-implanted devices," *Phys. Rev. Applied*, vol. 9, no. 2, p. 024033, Feb. 2018.
- [8] S. Gabriel, R. W. Lau, and C. Gabriel, "The dielectric properties of biological tissues: II. Measurements in the frequency range 10 Hz to 20 GHz," *Phys. Med. Biol.*, vol. 41, pp. 2251–2269, Nov. 1996.
- [9] D. Nikolayev, M. Zhadobov, P. Karban, and R. Sauleau, "Conformal antennas for miniature in-body devices: The quest to improve radiation performance," *URSI Rad. Sci. Bull.*, vol. 2017, no. 363, pp. 52–64, Dec. 2017.
- [10] A. Karlsson, "Physical limitations of antennas in a lossy medium," *IEEE Trans. Antennas Propag.*, vol. 52, no. 8, pp. 2027–2033, Aug. 2004.
- [11] D. Nikolayev, M. Zhadobov, and R. Sauleau, "Impact of tissue electromagnetic properties on radiation performance of in-body antennas," *IEEE Antenn. Wireless Propag. Lett.*, vol. 17, no. 8, pp. 1440–1444, Aug. 2018.
- [12] M. K. Magill, G. A. Conway, and W. G. Scanlon, "Tissue-independent implantable antenna for in-body communications at 2.36–2.5 GHz," *IEEE Trans. Antennas Propag.*, vol. 65, no. 9, pp. 4406–4417, Sep. 2017.
- [13] D. Nikolayev, M. Zhadobov, and R. Sauleau, "Immune-to-detuning wireless in-body platform for versatile biotelemetry applications," *IEEE Trans. Biomed. Circuits Syst.*, vol. 13, no. 2, pp. 403–412, Apr. 2019.
- [14] D. Nikolayev, W. Joseph, M. Zhadobov, R. Sauleau, and L. Martens, "Optimal radiation of body-implanted capsules," *Phys. Rev. Lett.*, vol. 122, no. 10, p. 108101, Mar. 2019.
- [15] Z. Bao, Y. X. Guo, and R. Mittra, "An ultrawideband conformal capsule antenna with stable impedance matching," *IEEE Trans. Antennas Propag.*, vol. 65, no. 10, pp. 5086–5094, Oct. 2017.
- [16] M. S. Miah, A. N. Khan, C. Icheln, K. Haneda, and K. Takizawa, "Antenna system design for improved wireless capsule endoscope links at 433 MHz," *IEEE Trans. Antennas Propag.*, vol. 67, no. 4, pp. 2687–2699, Apr. 2019.
- [17] R. S. Alrawashdeh, Y. Huang, M. Kod, and A. A. B. Sajak, "A broadband flexible implantable loop antenna with complementary split ring resonators," *IEEE Antenn. Wireless Propag. Lett.*, vol. 14, pp. 1506–1509, Feb. 2015.
- [18] D. Nikolayev, M. Zhadobov, L. Le Coq, P. Karban, and R. Sauleau, "Robust ultra-miniature capsule antenna for ingestible and implantable applications," *IEEE Trans. Antennas Propag.*, vol. 65, no. 11, pp. 6107–6119, Nov. 2017.
- [19] D. M. Pozar, *Microwave engineering*, 4th ed. Hoboken, NJ: Wiley, 2012.
- [20] S. Benaissa *et al.*, "Experimental characterization of in-to-out-body path loss at 433 MHz in dairy cows," *Electron. Lett.*, vol. 55, no. 7, pp. 422–424, Apr. 2019.